




CeraLux® Xenon Arc Lamps

Luxtel's CeraLux series of lamps provide the end market with a value supplier of ceramic lamps, wholly made in the United States. Originally engineered to be an exact replacement for lamps designed to operate in existing systems employing this type of ceramic, xenon short arc lamp. Today the series has been extended to provide custom designed lamps for major OEMs and includes collimated and focused, filtered and UV output in wattages from 150 to 500.

CeraLux lamps are designed to give the instrument builder a combination of rugged compactness and extremely high brightness, excellent color temperature over life and lamp to lamp repeatability.



CeraLux Xenon
Collimated Arc Lamp



CeraLux Xenon
Focused Arc Lamp


CeraLux xenon lamps produce broadband light with color temperatures of approximately 5900°K. This results in excellent white light for photopic, video and photographic applications.

LamCeraLux® Xenon Collimated Arc Lamps With Parabolic Reflector B Body Size



PHYSICAL SPECIFICATIONS

DESCRIPTION	SPECIFICATION
Reflector Geometry	Parabolic $Y^2 = .5X$
Window Diameter	1 inch (25.4 mm)
Weight	132 grams



Dimensions in parentheses are in millimeters

OPERATIONAL SPECIFICATIONS

	CL300BF	CL300BUV	CL175BF	CL175BUV	CL125BF	CL125BUV
POWER (watts)						
Nominal	300	300	175	175	125	125
Range	180 - 330	180 - 330	150 - 200	150 - 200	75 - 150	75 - 150
CURRENT (amps DC)						
Nominal	21	21	14	14	12	12
Range	10 - 22	10 - 22	12 - 16	12 - 16	8 - 14	8 - 14
VOLTAGE (volts DC)						
Nominal	14	14	12.5	12.5	11	11
Range	13 - 16	13 - 16	11 - 14	11 - 14	9.5 - 12.5	9.5 - 12.5
Minimum Ignition Voltage(kV)	23	23	23	23	23	23
Maximum Temperature	150°C	150°C	150°C	150°C	150°C	150°C

Typical Lifetime of lamps is 1000 hours – to 50% of initial output

OUTPUT AT NOMINAL POWER

Radiant Output (watts)	50	50	25	25	17	17
Visible Output (lumens)	5000	4500	2200	2000	1500	1350
UV Output (watts)	3	7	<2	4	<1	2
IR Output (watts)	29	27	14	13	10	9
Color Temperature (°K)	5800	5100	5800	5100	5800	5100
Cone Half Angle (at 10% pts)	5	5	5	5	5	5

All values and specifications subject to change without notice.

CeraLux[®] Xenon Arc Lamps

Read all safety, cooling, operational and disposal notes prior to use. Consult the factory at the numbers listed below for any assistance in the proper operation and use of all lamps.

SAFETY

- Always take precaution when handling lamps because they are under high pressure.
- The cold weld pinch-off tube is very sharp, care should be taken to not come into contact with it or disrupt it.
- During operation, the IR and UV radiation can cause skin and eye damage. Do not look directly into a lamp, get in the path of the output beam or touch the lamp while operating.
- Proper eye protection should be worn when handling lamps due to high pressure.
- Extra care should be exercised when using lamps that produce ultraviolet radiation.
- Protection of the eyes and skin should be undertaken when exposure to ultraviolet radiation is possible.
- When operating UV lamps, UV blocking eye protection should be worn during lamp operation.
- UV lamps will produce toxic ozone due to ultraviolet radiation below 242 nm acting upon oxygen in air. To avoid ozone, an inert atmosphere should be used. Select a non-UV lamp (filtered "F" series) if ultraviolet wavelengths are not needed.
- Due to high operating temperatures, care should be taken if handling lamps after operation.

COOLING

- Proper cooling will provide optimum performance and extend life. Forced air cooling should be used.
- Seal temperatures should not be allowed to exceed 200° C. Under nominal operating conditions attempts should be made to maintain maximum lamp temperatures of 150 ° C.
- Appropriate thermal and electrically conductive compounds should be used to promote improved thermal conduction between lamps and heat sinks. Do not apply thermal compounds to the lamp window or ceramic body. A thick layer should be applied to all surfaces that contact heat sinks.

OPERATION

- Lamp should not be operated with window facing upwards. Keep lamp oriented to within 45° of vertical.
- Lamps must be operated within stated current and voltage ranges. Overpowering can lead to electrode wear and shortened life. Underpowering may lead to instabilities and problems starting (depending on power source).
- Lamps should be mounted in suitable heat sinks.
- Assure proper polarity of power source when making electrical connections.
- Care should be taken when using filters or lenses that may reflect energy back into lamps.
- Consult the factory at the numbers below for proper use and operation of all lamps.

DISPOSAL

- Proper eye protection should be worn when handling lamps.
- Care must be taken due to the sharp pinch-off tube on the rear of the lamp.
- The internal pressure of the lamp may be relieved by opening the pinch-off tube with a pair of heavy pliers.
- The internal gas will be heard escaping through the tube opening. The lamp may now be disposed of.

WARRANTY

- LuxteL's CeraLux lamps are covered by a 500 hour warranty.
- Typical lifetime is 1000 hours (50% of initial output).



LuxteL[®] llc

100 Ferncroft Road ~ Danvers, MA 01923 USA

Tel. 978.774.9136 ~ Fax. 978.774.1559

Email: info@luxtel.net / Web: <http://www.luxtel.net>

msscscientific Chromatographie-Handel GmbH
Gneisenaustrasse 66/67 · 10961 Berlin · Germany
Fon: +49 30 6270 6087 · Fax: +49 30 6270 6089
Email: info@msscscientific.de · Web: www.msscscientific.de